

Original Research

Validity, Agreement and Reliability of the ForceFrame Dynamometer in Patients with Anterior Cruciate Ligament Injury

Kamilla Arp^{1,2a}, Thomas Frydendal^{3,4}, Troels Kjeldsen^{3,4}, Ulrik Dalgas⁵, Signe Timm⁶, Bjarke Viberg^{7,8}, Kim Ingwersen⁹, Claus Varnum^{2,6}

¹ Department of Orthopaedic Surgery, Lillebaelt Hospital, ² Department of Regional Health Research, University of Southern Denmark, ³ Department of Orthopedic Surgery, Aarhus University Hospital, ⁴ Department of Clinical Medicine, Aarhus University, ⁵ Exercise Biology, Department of Public Health, Aarhus University, ⁶ Department of Orthopedic Surgery, Lillebaelt Hospital, ⁷ Orthopaedic Surgery and Traumatology, Odense University Hospital, ⁸ Orthopaedic Surgery and Traumatology, Lillebaelt Hospital, ⁹ Department of Physio- and Occupational therapy, Lillebaelt Hospital

Keywords: ACL, knee, maximal muscle testing, validation

<https://doi.org/10.26603/001c.122486>

International Journal of Sports Physical Therapy

Vol. 19, Issue 9, 2024

Background

Restoring maximal muscle strength of the knee extensors (KE) and knee flexors (KF) following anterior cruciate ligament (ACL) injury and ACL reconstruction is of great importance to reduce the re-injury rate after ACL reconstruction and to reduce the risk of knee osteoarthritis. Therefore, it is essential that clinicians and healthcare providers use valid and reliable measures to assess knee muscle strength to ensure a safe return to sport.

Purpose

To evaluate the reliability (test-retest reliability, inter-tester reliability and test-retest agreement) and validity (concurrent validity, convergent validity and ForceFrame (FF) vs. isokinetic dynamometer (ID) agreement) of the ForceFrame (FF) dynamometer during isometric testing of the knee extensors and flexors.

Study Design

Cross-sectional study

Material and Methods

Twenty-seven participants with ACL injury or reconstruction were recruited for participation in this study. maximal voluntary isometric contraction (MVIC) of the knee extensors and flexors was tested on two separate days. Day one included validity assessments with FF, a gold-standard ID and a handheld dynamometer (HHD). Day two included reliability assessments with FF performed by two assessors. Main outcome measures were day-to-day test-retest reliability and agreement and inter-tester reliability of FF, and concurrent validity (FF vs. an ID and a HHD). Reliability was tested as test-retest and inter-tester reliability using interclass correlation coefficient (ICC), while agreement was tested using Bland & Altman plots with limits of agreement (LOA), standard error of measurement (SEM) and smallest detectable change (SDC). Concurrent validity between FF, ID, and HHD was assessed using Pearson's correlations and mean difference was evaluated by Bland & Altman plots.

Results

Twenty-seven participants (10 females, 17 males) with a median age of 25 years (range 19-60) were included in this study. There was a good day-to-day test-retest reliability for

^a Corresponding author:

Kamilla Arp

Address: Lillebaelt Hospital, Beriderbakken 4, 7100 Vejle, Denmark

E-mail: kamilla.dessau.arp@rsyd.dk

Tel: +457940 9132

MVIC of KE (ICC=0.77, CI95:0.48-0.90) and KF (ICC=0.83, CI95:0.61-0.92) and excellent inter-tester reliability for MVIC of KE (ICC=0.97, CI95:0.94-0.98) and KF (ICC=0.93, 95CI:0.85-0.97). Standard error of measurement (SEM) was 8% and 9%, while the smallest detectable change (SDC) was 22% and 27% for KE and KF, respectively. FF showed fair concurrent validity compared to ID for KE ($r=0.56$), poor concurrent validity for knee flexors (KF ($r=0.24$) and compared to HHD a moderate correlation for KE ($r=0.74$) and poor correlation for KF ($r=0.12$). Bland & Altman plots between FF and the ID showed a mean difference of -0.51 Nm/kg for KE and -0.32Nm/kg for KF.

Conclusions

FF can be used to obtain reliable and valid results to assess MVIC of the KE, but not the KF. It should be noted that absolute results produced by the FF may be considered an underestimation of actual MVIC. The test position to assess KF in FF does not appear to be optimal, and different test-positions may be considered.

Level of evidence

Level 3

INTRODUCTION

Anterior cruciate ligament (ACL) injury is the most common ligamentous knee injury.¹ ACL injuries frequently occur during sports² and are typically associated with muscle weakness, risk of instability and poor knee function.^{3,4} These impairments are usually present in the short term but often persist long term after ACL injury.⁵ Symmetrical knee extensor (KE) strength reduces the re-injury rate after ACL reconstruction (ACLR)⁶ and higher KE strength early after ACL injury predicts improved patient reported and functional performance outcomes after ACLR.^{7,8} Patients with ACL injuries have an increased risk of osteoarthritis (OA) as compared to the general population,⁹ but strengthening KE muscles plays a pivotal role in preventing knee OA.¹⁰ Further, knee flexor (KF) strength plays an important role in protecting the ACL.¹¹ KF strength is particularly decreased after ACL reconstruction with hamstring graft.¹² This may persist more than two years after ACL reconstruction^{13,14} putting the patient at risk of a secondary ACL injury. Taken together, restoring maximal knee muscle strength and limb symmetry index (LSI) following ACL injury and ACLR is therefore of great importance.

Considering the importance of evaluating muscle strength following an ACL injury, it is essential that assessment methods are easy, reliable, and valid. Assessment of maximal voluntary isometric muscle contraction (MVIC) using an isokinetic dynamometer (ID) has previously shown excellent clinimetric properties and is considered the “gold standard”.¹⁵ However, IDs are expensive, require training to handle, and are not portable, making them unavailable for most clinicians.

Handheld dynamometers (HHD) are affordable, easy to use, portable and frequently used to evaluate MVIC in clinical practice. Although HHD appear to have acceptable clinimetrics when evaluating KE strength after ACLR,¹⁶ they may underestimate the force output at higher force values,¹⁷ and it can be difficult to ensure a reliable setup.¹⁸

The ForceFrame (FF) was recently developed for assessment of maximal muscle strength. The FF (Figure 1) has foldable plates, an adjustable crossbar (which can be ad-




Figure 1. The ForceFrame dynamometer

justed in height and can be rotated) and the force transducers are sensors that can also be adjusted and rotated. The FF is portable and does not require extensive training, making it easy to use for clinicians.

Previous studies have evaluated the FF for test-retest reliability when assessing hip strength^{19,20} and strength of the shoulder rotators.²¹ These showed acceptable reliability for both hip strength (in a non-injured athletic population) and shoulder strength (in healthy athletes). However, the clinimetric properties of the FF to assess MVIC of the KE and KF in patients with ACL injuries have not previously been evaluated. Therefore, the purpose of this study was to evaluate the reliability (test-retest reliability, inter-tester reliability and test-retest agreement) and validity (concurrent validity, convergent validity and FF vs. ID agreement) of the FF dynamometer during isometric testing of the knee extensors and flexors.

MATERIALS AND METHODS

STUDY DESIGN AND PARTICIPANTS

This observational study was performed at the Department of Public Health, Aarhus University, Denmark and evaluated participants with an ACL injury or an ACL reconstruction. Participants were recruited from a public rehabilita-

tion centre in Aarhus, Denmark between April and October 2023. Inclusion criteria were: 1) participants having had an ACL injury or ACLR in the prior 6-24 months, 2) ≥ 18 years of age and 3) able to speak, read and understand Danish. All participants received oral and written information regarding participation in the study and gave informed oral consent. The study is registered at the region of Southern Denmark.

EXPERIMENTAL SET-UP

All participants completed two test sessions one to two weeks apart. During the first test session participants completed: 1) a background questionnaire, 2) testing of MVIC of the KE and KF using both FF (Vald Performance, Queensland, Australia), an HHD (Commander Echo Wireless Console and Muscle Tester, JTECH Medical, Salt Lake City, Utah, USA) and an ID (Humac NORM, CSMi, MA, USA), 3) registration of self-reported pain (using a numerical rating scale [NRS]) in the knee being tested before and after each test direction in all three test devices and 4) the "Knee Outcome Survey – Activities of Daily Living Scale" (KOS-ADLS).²² At the second test session participants completed 1) two tests of MVIC of the KE and KF using only the FF performed by two different assessors and 2) a battery of single legged hop tests if they had a LSI between the injured and uninjured leg $\geq 90\%$ for the KE.

To minimize bias in the test results, it was decided that the order of testing should be randomized. In the first test session (with FF, ID, and HHD), three patients could be tested simultaneously. On some test days, only one or two patients were scheduled simultaneously. The order of testing for each patient was randomized by a simple draw (three pieces of paper indicating the three devices) to determine which device would be tested first. In the second session, only one patient at a time could be tested, but was tested by two different assessors. The draw decided which assessor would perform the first assessment.

ASSESSORS

Three assessors (7-13 years of experience) who were trained with the use of the FF, ID or HHD participated in this study. In the first session, two physiotherapists (KA, TF) and one exercise physiologist (TK) performed the tests with the FF, HHD and ID, respectively. At the second session, only two assessors were needed (KA, TF) to test all the participants using the FF.

TESTING PROCEDURES

MVIC

Before the MVIC test of KE and KF, a mark was made on the skin approximately 5 cm proximal to the lateral malleolus to guide the placement of the force transducer from the HHD and FF. However, both assessors (in HHD and FF) independently measured the lever arm from the lateral joint line in the knee to the skin mark. In the ID, the lever arm was measured from the placement of the lever arm on the




Figure 2 A and B. Test position during the ForceFrame assessment while performing maximal voluntary isometric contractions of the knee extensors (A) and knee flexors (B)

device and no mark on the skin was made for this test. After measuring the lever arm participants performed a standardized five minute warm up on a stationary bike. The warm-up load was adjusted according to the patients' rated perceived exertion (RPE). During the first three minutes, patients aimed for an RPE of 5-6 (out of 10). In the final two minutes, the target exertion level was increased to an RPE of 8-9. Tests of KE (regardless of device) were performed with the participant sitting on a bench with 90 degrees flexion in the hip and the lower leg hanging down with the knee in approximately 90 degrees flexion (Figures 2A, 2C and 2E).

Tests of KF (regardless of device) were performed with the participant seated upright with a knee angle of approximately 70 degrees (Figures 2B, 2D and 2F) and the force transducer placed on the lower part of the calf guided by the same mark as used for KE testing.

For the FF, a tailor-made bar was used to stabilize the device during the assessment of KE (Figure 2A). For the HHD, the device was fastened using a custom-made belt for stabilization during both KE and KF assessments (Figures 2C and 2D). For the ID, participants had a back support and were fastened using a belt over the waist (Figures 2E and 2F).

Prior to starting a test on each device, a standardized test-specific warm up protocol was performed as previously described.²³ This consisted of three submaximal trials where the participants were instructed to contract at approximately 50%, 75% and 90% of their perceived maximal force. After this, at least three maximal trials were completed, separated by 45 seconds of rest between trials. If the third trial resulted in the highest value, participants were given extra attempts until no improvement occurred.

For each trial, participants were instructed to sit upright with the hands across the chest and to relax the leg being tested, while the assessor counted down from three to start the test. Participants were further instructed to press 'as hard and fast as possible', while trying to maintain the pres-




Figure 2 C and D. Test position of the hand-held dynamometer assessment while performing maximal voluntary isometric contractions of the knee extensors (C) and knee flexors (D)




Figure 2 E and F. Test position during the isokinetic dynamometer (Humac Norm) assessment of maximal voluntary isometric contractions of the knee extensors (E) and knee flexors (F)

sure for three seconds. Standardized verbal encouragement was provided during each trial.

For both KE and KF, the non-injured limb was tested first and the best trial for KE and KF from both limbs were used for the analyses and reported as Nm/kg.

NUMERAL RANG SCALE (NRS)

Before and after each trial for both KE and KF, participants rated their current knee pain on a 11 point NRS ranging from 0 (no pain) to 10 (worst imaginable pain).²⁴ This was collected to evaluate if knee pain increased following the tests to ensure that no unacceptable pain occurred. The test session was terminated if participants rated their knee pain above 5 on the NRS.

KNEE OUTCOME SURVEY-ACTIVITIES OF DAILY LIVING (KOS-ADLS)

KOS-ADLS is a reliable, valid and responsive knee-specific patient reported outcome measure (PROM) measuring

symptoms and functional limitations with activities of daily living ranging from 0 (worst) to 100 (best).^{22,25} Participants completed the KOS-ADLS at the end of the first test-session.

FUNCTIONAL PERFORMANCE

Functional performance was assessed using a battery of four single legged hop tests: 1) single-leg-hop for distance (SLHD), 2) cross-over-hop for distance (CHD), 3) triple-hop for distance (THD) and 4) 6-m timed hop (6m-TH), performed as described previously.²⁶ Results for each single leg hop test were reported as an average of two trials and presented in centimeters (SLHD, CHD and THD) or seconds (6m-TH).

Only six participants completed the battery of single legged hop tests and these results are therefore not included in this paper.

STATISTICAL ANALYSIS

Descriptive statistics were used to describe participant characteristics. Dichotomous and ordinal data are presented as frequencies (%) and continuous data are presented as means and SD if data is normally distributed or median and range if data is not normally distributed.

Prior to any analysis, QQ plots and histograms were visually inspected to ensure normal distribution of data. All results were presented for the KE and KF for both the injured leg and the non-injured leg.

RELIABILITY

TEST-RETEST RELIABILITY

Day-to-day test-retest and inter-tester reliability was analyzed using interclass correlation coefficient (ICC) and reported with 95% confidence intervals (CI), based on single measures, absolute agreement and two-way mixed effects. Reliability was interpreted as poor (ICC < 0.50), moderate (ICC 0.50-0.74), good (ICC between 0.75-0.90) or excellent (ICC > 0.90).²⁷ Day-to-day test-retest reliability was calculated from the best trial for both KE and KF from the first and second FF test-sessions (same test-leader). Inter-tester reliability was calculated from the best trial for both KE and KF of the second test-session (two assessors).

DAY-TO-DAY TEST-RETEST AGREEMENT

To evaluate day-to-day test-retest agreement, the best trials for KE and KF from the first and second FF test-session (same test-leader) were used. The mean difference in force (Nm/kg) between test and re-test was plotted against the mean force of the two measurements with 95% limits of agreement (LOA). LOA was calculated as the standard deviation (SD) of the differences between the day-to-day tests multiplied by 1.96. The standard error of measurement (SEM) was used to assess the absolute error of the measures and the smallest detectable change (SDC) was calculated (using the squareroot of the error variance) to reflect the

smallest change exceeding test variation. SEM and SDC were calculated and presented as percentages of the group mean force.

VALIDITY

Concurrent validity was analyzed (Pearsons correlation) to evaluate the relationship between the muscle strength results measured using FF and ID and between FF and HHD. Correlation was interpreted as poor ($r < 0.30$), fair ($r = 0.30-0.59$), moderate ($0.60-0.79$) and very strong ($r > 0.80$).^{28,29} In addition, we also analyzed agreement between assessments with the FF and the ID to report the mean differences.

Convergent validity evaluated the relationship (Spearman correlation) between muscle strength measures and the outcome on the KOS-ADLS.

RESULTS

In total, 24 of 27 completed both test sessions. Of the three that did not complete the second test session, one had increased knee pain from the first test session that did not resolve before the second session, one failed to attend the scheduled assessment, and one was prevented due to personal reasons. Of the 27 participants, 10 were female and 17 were male. The median age of the participants was 25 (19-60) years (Table 1). Mean MVIC values for KE in the injured leg were $1.80 (\pm 0.32)$ Nm/kg in FF, $2.31 (\pm 0.50)$ Nm/kg in the ID and $1.68 (\pm 0.40)$ Nm/kg in the HHD. For KF, mean MVIC values in the injured leg were $0.85 (\pm 0.22)$ Nm/kg in the FF, $1.18 (\pm 0.33)$ Nm/kg in the ID and $1.07 (\pm 0.26)$ Nm/kg in the HHD (Table 2).

RELIABILITY

TEST-RETEST RELIABILITY

The FF presented with good day-to-day test-retest reliability for measuring MVIC of KE (ICC = 0.77, 95% CI 0.48-0.90) and KF (ICC = 0.83, 95% CI 0.61-0.92) in the injured leg as well as good day-to-day test-retest reliability of KE (ICC = 0.80, 95% CI 0.56-0.91) and KF (ICC = 0.87, 95% CI 0.72-0.94) in the non-injured leg (Figure 3A). The inter-tester reliability was excellent for measuring MVIC of KE (ICC = 0.97, 95% CI 0.94-0.98) and KF (ICC = 0.93, 95% CI 0.85-0.97) in the injured leg, and a good inter-tester reliability was observed for KE (ICC = 0.92, 95% CI 0.91-0.98) and KF (ICC = 0.91, 95% CI 0.64-0.97) of the non-injured leg (Figure 3B).

DAY-TO-DAY TEST-RETEST AGREEMENT

Bland & Altman plots shows a wide LOA between test and retest for KE and KF for both the injured and noninjured leg (Figures 4A, 4B and 5A and 5B), but with homogeneous distribution of datapoints. For the injured leg, KE showed LOA ranging from -0.57 to 0.51 Nm/kg and for KF LOA ranged from -0.29 to 0.36 (figure 4A and 5A). The SEM for KE and KF on the injured leg was 8% and 9%, respectively. The SDC

for KE and KF was 22% and 27% for the injured leg, respectively (Table 3).

VALIDITY

Concurrent validity between FF and ID showed a fair correlation for KE ($r = 0.56$) and a poor correlation for KF ($r = 0.24$) when evaluated in the injured leg, and a moderate correlation for KE ($r = 0.65$) and a fair correlation for KF ($r = 0.36$) when evaluated in the non-injured leg (Figure 6 A). In the injured leg, a moderate correlation was found between FF and HHD for knee extension (KE) ($r = 0.74$) and a poor correlation for KF ($r = 0.12$) (Figure 6B). In the non-injured leg, there was a moderate correlation for KE ($r = 0.66$) and a fair correlation for KF ($r = 0.49$) (Figure 6B). Results from the injured leg from all devices are visually presented for KE in Figure 7A and KF in Figure 7B. Bland & Altman plot between FF and ID for KE on injured leg showed a mean difference of -0.51 Nm/kg and for KF on injured leg a mean difference of -0.32 Nm/kg (Figure 8 A and B).

Convergent validity between FF and the KOS-ADLS was found to be poor for KE ($r = 0.13$) and negative for KF ($r = -0.11$).

DISCUSSION

This is the first study to evaluate the reliability and validity of the FF for assessing MVIC of KE and KF in an ACL injured population. The FF showed: 1) good day-to-day test-retest reliability, when assessing MVIC of KE and KF for the injured leg and excellent inter-tester reliability when assessing MVIC of KE and KF, 2) that a change exceeding 22-27% is needed in the evaluation KE and KF on the injured leg to ensure that a true change has taken place, 3) a moderate concurrent validity when assessing KE and a poor concurrent validity when assessing KF on the injured leg compared to the ID and 4) poor convergent validity as neither KE nor KF strength on the injured leg were associated with the score on KOS-ADLS.

RELIABILITY

Good day-to-day test-retest reliability and excellent inter-tester reliability were found for both KE and KF on both the injured and non-injured leg when using the FF. This suggests that assessments on FF can be reproduced between test days and that assessments are not impacted markedly by different assessors. The high ICCs observed are likely enforced by the highly standardized test protocol, where participants were given standardized instructions during all assessments. On the second test-day, participants were asked if they experienced a change in their knee problems since the first test-day. Four participants (of 24) reported that they experienced a change in their knee problems (two better, two worse), which might have impacted their results. Furthermore, it was not possible to ensure that assessments were completed on the same time of day, which therefore allow some diurnal variation to have occurred. However, the two assessments evaluating inter-tester reliability

Table 1. Characteristics of the participants (n = 27)

Variable	
Females, n (%)	10 (37)
Males, n (%)	17 (63)
Age (years)*	26 (19-60)
BMI (kg/m ²)*	25 (19-34)
Primary ACL rupture, n (%)	23 (85)
Secondary ACL rupture, n (%)	3 (11)
Previous contralateral ACL rupture, n (%)	1 (4)
Current use of pain medication, yes (%)	1 (4)
KOS-ADLS, 0-100*	85 (44-99)
Injury during sports, n (%)	25 (92)
Pre-injury sports active, yes (%)	26 (96)
Pre-injury activity frequency (n = 26)	
1 x/week, n (%)	2 (7)
2-3 x/week, n (%)	14 (53)
4-5 x/week, n (%)	9 (34)
6 or more x/week, n (%)	1 (3)
Pre-injury activity level (n = 26)	
Recreational (non-competitive), n (%)	7 (26)
Recreational (competitive), n (%)	14 (53)
Elite (national elite level)**, n (%)	5 (19)
Returned to sports (n = 26), yes (%)	11 (40)
Returned to pre-injury sport (n = 11), yes (%)	3 (27)
Returned to pre-injury level of sport (n=3), yes (%)	1 (33)
Rehabilitation only, n (%)	17 (26)
ACL reconstruction, n (%)	20 (74)
Graft-type (n = 20)	
Hamstrings graft, n (%)	12 (60)
Quadriceps graft, n (%)	7 (35)
Patellar tendon (BPTB), n (%)	0
Allograft, n (%)	1 (5)
Time from surgery to test (n=20), months*	9 (6-12)
Time from injury to test (rehab only) (n=7), months*	9 (6-11)

***Reported as median (range). BMI body mass index, ACL anterior cruciate ligament, KOS ADLS Knee Outcome Survey - Activities of Daily Living ranging from 0 (worst) to 100 (best)

** 1st to 3rd national division

Table 2. Mean MVIC in KE and KF on injured and non-injured leg

	Knee extension		Knee flexion	
	Injured leg	Non-injured leg	Injured leg	Non-injured leg
ForceFrame	1.80 (±0.32) Nm/kg	2.01 (±0.30) Nm/kg	0.85 (±0.22) Nm/kg	0.75 (±0.21) Nm/kg
ID	2.31 (±0.50) Nm/kg	2.89 (±0.67) Nm/kg	1.18 (±0.33) Nm/kg	1.30 (±0.30) Nm/kg
HHD	1.68 (±0.40) Nm/kg	1.99 (±0.50) Nm/kg	1.07 (±0.26) Nm/kg	1.20 (±0.28) Nm/kg

MVIC; maximal voluntary isometric contraction, KE; knee extension, KF; knee flexion, ID; isokinetic dynamometer, HHD; hand held dynamometer

bility were performed on the same day, therefore excluding day-to-day variation.

Although no previous studies have evaluated reliability of the FF for KE, the FF has shown good test-retest reliability (ICC = 0.77-0.95) when assessing hip strength in football

players²⁰ and excellent test-retest reliability in Australian footballers (ICC = 0.87-0.97).³⁰ Similar test-retest reliability results were found in a study evaluating shoulder rotational strength (ICC = 0.85-0.92), which may indicate that the FF




Figure 3 A and B. Interclass correlation coefficient (ICC) with 95% confidence intervals for test-retest (A) and inter-tester (B) reliability




Figure 4 A and B. Bland-Altman plots for the ForceFrame MVIC knee extension of the injured leg (A) and non-injured leg (B)




Figure 5 A and B. Bland-Altman plots for the ForceFrame MVIC knee flexion of the injured leg (A) and non-injured leg (B)

is a reliable device for evaluating MVIC and not sensitive to different assessors when specific protocols are applied.

AGREEMENT

LOA ranged from -0.57 to 0.51 Nm/kg for KE and from -0.29 to 0.36 Nm/kg for KF. These ranges are quite wide consider-

Table 3. SEM and SDC calculations relative to mean force produced in the ForceFrame

	SEM	SDC
Knee extension- injured	8%	22%
Knee extension – non-injured	6%	19%
Knee flexion – injured	9%	27%
Knee flexion - non-injured	8%	24%

SEM; standard error of measurement, SDC; smallest detectable change




Figure 6 A and B. Correlation coefficients (rho) for concurrent validity between the ForceFrame and the isokinetic dynamometer (A) and the ForceFrame and the handheld dynamometer (B)




Figure 7 A and B. Mean force outputs during MVIC testing using the ForceFrame, the isokinetic dynamometer and the handheld dynamometer for knee extension (A) and knee flexion (B)




Figure 8 A and B. Bland-Altman plots for MVIC assessed with the ForceFrame and the isokinetic dynamometer for knee extension (A) and knee flexion (B)

ing the mean force of 1.80 Nm/kg and for KE and 0.85 Nm/kg for KF. The mean difference between the two test-days were only -0.03 Nm/kg for KE and 0.03 Nm/kg for KF, but considering the wide LOA, there were large differences in the absolute test results between days. This could be explained by a difference in the knee condition between test-days, but is most likely related to a less stable test position compared to the ID. A previous study evaluating agreement between HHD and an ID reported no agreement between devices when evaluating KE MVIC³¹ and discussed the impact of a stable test position, which likely also had an impact on the results in the present study.

The evaluation of agreement between the FF and ID indicated that FF measured 0.50 Nm/kg and 0.32 Nm/kg lower than the ID in KE and KF. As agreement evaluates absolute values it must be considered that FF may not express the true MVIC considering the underestimation of KE and KF strength when using the FF. The difference between devices are again most likely explained by the different test positions, where the ID uses more optimal positioning and stabilization, allowing the participants to produce higher force outputs, as also suggested by previous research.³² From %SDC it is also notable that a change of less than 22% in KE and 27% in KF may reflect a measurement error rather than a real change. A previous review evaluated HHDs compared to IDs for assessing MVIC KE and KF and found that LOA were generally higher than 10%.³² Therefore, clinicians must carefully consider the magnitude of change over time when they interpret whether a strength improvement reflect an actual improvement rather than a potential measurement error. Of note, agreement was comparable for KE between both the injured and non-injured leg, suggesting that measurement error is not only reflected on the injured leg.

VALIDITY

A fair and moderate concurrent validity between FF and the ID for KE was found for both the injured and the non-injured leg, whereas a poor and fair concurrent validity was found for KF for both the injured and non-injured leg. This

indicates that the FF can be used to establish a valid MVIC of KE, whereas KF assessments should be interpreted very cautiously. The concurrent validity between FF and HHD for KE was higher than between FF and the ID, which may be explained by more comparable setups between these devices, as compared to the ID. When tested in FF and HHD, participants were seated on a bench without belt fixation over the hip nor had any backrest support, which is contrary to the ID. The weaker correlation between FF and ID than between FF and HHD is likely explained by the more optimal conditions available for force production in the ID, as FF generated 22% less force than the ID and only 7% more force than the HHD. The ID setup has the advantage of a completely fixated lever arm and seat, cushioning on the lever arm and seat to avoid pain, and requires less attention paid towards maintaining a stable position in the upper body and to keep seated when applying maximal force.³¹

Previous studies have evaluated concurrent validity for the HHD using ID as reference,^{31,33-36} reporting moderate to excellent correlations. Although not directly comparable, this indicates that KE FF validity is within the expected range. The poor correlations between the FF and ID for KF on the injured leg suggest that the applied test position might not be optimal for assessing KF muscle strength. In this group of patients with ACL injury, 12 participants had ACL reconstruction with hamstring graft, which are known to experience difficulties in regaining KF strength despite rehabilitation³⁷ and could possibly influence the assessments if the setup for KF is more challenging than in the ID. In the present study, patients with hamstrings graft had similar KF strength in FF compared to patients with other grafts or rehabilitation only, but lower KF strength in the ID compared to the rest of the group. Furthermore, a recent study found that deficits in KF strength for patients with ACL reconstruction was more accurately evaluated using a device assessing eccentric KF in the Nordic hamstrings exercise compared to an ID.³⁸ Given the various adjustment options in FF it may be considered if other test positions should be adopted when evaluating KF, but this remains to be investigated.

METHODOLOGICAL CONSIDERATIONS

This study had some limitations that should be considered when interpreting the results. First, as this study evaluated reliability, validity, and agreement of KE and KF with FF in patients with ACL injuries, the results are limited to this group of patients. Second, the authors used a tailor-made bar to stabilize the FF in the assessment of KE. This must be considered since this stabilizing bar is not part of the commercial FF kit, but was deemed necessary to ensure sufficient stability during the KE assessment. Third, despite using a standardized protocol, it was based on a subjective evaluation if an attempt was not accepted. A higher force output is likely occurring if a participant did not maintain seated or excessively pulled the upper body back. However, it is also possible that participants might feel limited in producing maximal force by the restrictions in the standardized test position.

CONCLUSIONS

The results of this study indicate that the FF can be used to obtain reliable and valid assessments of MVIC of KE. However, absolute values must be interpreted with great caution as FF likely underestimates the true MVIC value compared to gold standard (ID) and changes must exceed 22-27% to be considered a real change. If the FF assessments are used for evaluative and decision-making purposes this must be taken into account. The test position to assess KF in FF does not appear to be optimal, and different test-positions should be considered.

.....

DISCLOSURES

The study was supported with funding (to purchase of a test device, the ForceFrame and covering transportations costs for participants) by the Research Council at Lillebaelt Hospital, Denmark. The authors report no conflicts of interest associated with the creation of this manuscript.

ACKNOWLEDGEMENTS

Great thanks to project nurses Jane Leonhardt and Pia Hostrup Andersen for assisting in practical planning, preparing and coordinating the test sessions

Submitted: February 22, 2024 CDT, Accepted: July 19, 2024

CDT

© The Author(s)



This is an open-access article distributed under the terms of the Creative Commons Attribution 4.0 International License (CCBY-NC-4.0). View this license's legal deed at <https://creativecommons.org/licenses/by-nc/4.0> and legal code at <https://creativecommons.org/licenses/by-nc/4.0/legalcode> for more information.

REFERENCES

1. Mall NA, Chalmers PN, Moric M, et al. Incidence and trends of anterior cruciate ligament reconstruction in the United States. *Am J Sports Med.* 2014;42(10):2363-2370. doi:10.1177/0363546514542796
2. Granan LP, Forssblad M, Lind M, Engebretsen L. The Scandinavian ACL registries 2004-2007: Baseline epidemiology. *Acta Orthop.* 2009;80(5):563-567. doi:10.3109/17453670903350107
3. Tayfur B, Charupongsang C, Morrissey D, Miller SC. Neuromuscular function of the knee joint following knee injuries: Does it ever get back to normal? A systematic review with meta-analyses. *Sports Med.* 2021;51(2):321-338. doi:10.1007/s40279-020-01386-6
4. Lisee C, Lepley AS, Birchmeier T, O'Hagan K, Kuenze C. Quadriceps strength and volitional activation after anterior cruciate ligament reconstruction: A systematic review and meta-analysis. *Sports Health.* 2019;11(2):163-179. doi:10.1177/1941738118822739
5. Filbay SR, Culvenor AG, Ackerman IN, Russell TG, Crossley KM. Quality of life in anterior cruciate ligament-deficient individuals: A systematic review and meta-analysis. *Br J Sports Med.* 2015;49(16):1033-1041. doi:10.1136/bjsports-2015-094864
6. Grindem H, Snyder-Mackler L, Moksnes H, Engebretsen L, Risberg MA. Simple decision rules can reduce reinjury risk by 84% after ACL reconstruction: The Delaware-Oslo ACL cohort study. *Br J Sports Med.* 2016;50(13):804-808. doi:10.1136/bjsports-2016-096031
7. Eitzen I, Holm I, Risberg MA. Preoperative quadriceps strength is a significant predictor of knee function two years after anterior cruciate ligament reconstruction. *Br J Sports Med.* 2009;43(5):371-376. doi:10.1136/bjsm.2008.057059
8. Lepley LK, Palmieri-Smith RM. Quadriceps strength, muscle activation failure, and patient-reported function at the time of return to activity in patients following anterior cruciate ligament reconstruction: A cross-sectional study. *J Orthop Sports Phys Ther.* 2015;45(12):1017-1025. doi:10.2519/jospt.2015.5753
9. Oiestad BE, Juhl CB, Culvenor AG, Berg B, Thorlund JB. Knee extensor muscle weakness is a risk factor for the development of knee osteoarthritis: An updated systematic review and meta-analysis including 46 819 men and women. *Br J Sports Med.* 2022;56(6):349-355. doi:10.1136/bjsports-2021-104861
10. Whittaker JL, Losciale JM, Juhl CB, et al. Risk factors for knee osteoarthritis after traumatic knee injury: A systematic review and meta-analysis of randomised controlled trials and cohort studies for the OPTIKNEE Consensus. *Br J Sports Med.* 2022;56(24):1406-1421. doi:10.1136/bjsports-2022-105496
11. Bencke J, Aagaard P, Zebis MK. Muscle activation during ACL injury risk movements in young female athletes: A narrative review. *Front Physiol.* 2018;9:445. doi:10.3389/fphys.2018.00445
12. Sinding KS, Nielsen TG, Hvid LG, Lind M, Dalgas U. Effects of autograft types on muscle strength and functional capacity in patients having anterior cruciate ligament reconstruction: A randomized controlled trial. *Sports Med.* 2020;50(7):1393-1403. doi:10.1007/s40279-020-01276-x
13. Bregenhof B, Aagaard P, Nissen N, et al. The effect of progressive resistance exercise on knee muscle strength and function in participants with persistent hamstring deficit following ACL reconstruction: A randomized controlled trial. *J Orthop Sports Phys Ther.* 2023;53(1):40-48. doi:10.2519/jospt.2022.11360
14. Heijne A, Werner S. A 2-year follow-up of rehabilitation after ACL reconstruction using patellar tendon or hamstring tendon grafts: A prospective randomised outcome study. *Knee Surg Sports Traumatol Arthrosc.* 2010;18(6):805-813. doi:10.1007/s00167-009-0961-3
15. de Araujo Ribeiro Alvares JB, Rodrigues R, de Azevedo Franke R, et al. Inter-machine reliability of the Biodex and Cybex isokinetic dynamometers for knee flexor/extensor isometric, concentric and eccentric tests. *Phys Ther Sport.* 2015;16(1):59-65. doi:10.1016/j.ptsp.2014.04.004
16. Almeida GPL, Albano TR, Melo AKP. Hand-held dynamometer identifies asymmetries in torque of the quadriceps muscle after anterior cruciate ligament reconstruction. *Knee Surg Sports Traumatol Arthrosc.* 2019;27(8):2494-2501. doi:10.1007/s00167-018-5245-3

17. Kittelson AJ, Christensen JC, Loyd BJ, Burrows KL, Iannitto J, Stevens-Lapsley JE. Reliability, responsiveness, and validity of handheld dynamometry for assessing quadriceps strength in total knee arthroplasty. *Disabil Rehabil*. 2021;43(21):3070-3077. doi:10.1080/09638288.2020.1730454
18. Hansen EM, McCartney CN, Sweeney RS, Palimenio MR, Grindstaff TL. Hand-held dynamometer positioning impacts discomfort during quadriceps strength testing: A validity and reliability study. *Int J Sports Phys Ther*. 2015;10(1):62-68.
19. O'Brien M, Bourne M, Heerey J, Timmins RG, Pizzari T. A novel device to assess hip strength: Concurrent validity and normative values in male athletes. *Phys Ther Sport*. 2019;35:63-68. doi:10.1016/j.pts.2018.11.006
20. Desmyttere G, Gaudet S, Begon M. Test-retest reliability of a hip strength assessment system in varsity soccer players. *Phys Ther Sport*. 2019;37:138-143. doi:10.1016/j.pts.2019.03.013
21. Couch J, Sayers M, Pizzari T. Reliability of the ForceFrame with and without a fixed upper-limb mold in shoulder rotation strength assessments compared with traditional hand-held dynamometry. *J Sport Rehabil*. 2021;30(8):1246-1249. doi:10.1123/jsr.2020-0434
22. Irrgang JJ, Snyder-Mackler L, Wainner RS, Fu FH, Harner CD. Development of a patient-reported measure of function of the knee. *J Bone Joint Surg Am*. 1998;80(8):1132-1145. doi:10.2106/00004623-199808000-00006
23. Aagaard P, Simonsen EB, Andersen JL, Magnusson P, Dyhre-Poulsen P. Increased rate of force development and neural drive of human skeletal muscle following resistance training. *J Appl Physiol* (1985). 2002;93(4):1318-1326. doi:10.1152/japplphysiol.00283.2002
24. Hawker GA, Mian S, Kendzerska T, French M. Measures of adult pain: Visual Analog Scale for Pain (VAS Pain), Numeric Rating Scale for Pain (NRS Pain), McGill Pain Questionnaire (MPQ), Short-Form McGill Pain Questionnaire (SF-MPQ), Chronic Pain Grade Scale (CPGS), Short Form-36 Bodily Pain Scale (SF-36 BPS), and Measure of Intermittent and Constant Osteoarthritis Pain (ICOAP). *Arthritis Care Res (Hoboken)*. 2011;63 Suppl 11:S240-52. doi:10.1002/acr.20543
25. Irrgang JJ, Anderson AF. Development and validation of health-related quality of life measures for the knee. *Clin Orthop Relat Res*. 2002;(402)(200209000-00009):95-109. doi:10.1097/00003086-200209000-00009
26. Grindem H, Eitzen I, Engebretsen L, Snyder-Mackler L, Risberg MA. Nonsurgical or surgical treatment of ACL Injuries: Knee function, sports participation, and knee reinjury: The Delaware-Oslo ACL cohort study. *J Bone Joint Surg Am*. 2014;96(15):1233-1241. doi:10.2106/BJS.M.01054
27. Koo TK, Li MY. A guideline of selecting and reporting intraclass correlation coefficients for reliability research. *J Chiropr Med*. 2016;15(2):155-163. doi:10.1016/j.jcm.2016.02.012
28. Akoglu H. User's guide to correlation coefficients. *Turk J Emerg Med*. 2018;18(3):91-93. doi:10.1016/j.tjem.2018.08.001
29. Chan YH. Biostatistics 104: Correlational analysis. *Singapore Med J*. 2003;44(12):614-619.
30. Ryan S, Kempton T, Pacecca E, Coutts AJ. Measurement properties of an adductor strength-assessment system in professional Australian footballers. *Int J Sports Physiol Perform*. 2019;14(2):256-259. doi:10.1123/ijsp.2018-0264
31. Martins J, da Silva JR, da Silva MRB, Bevilacqua-Grossi D. Reliability and validity of the belt-stabilized handheld dynamometer in hip- and knee-strength tests. *J Athl Train*. 2017;52(9):809-819. doi:10.4085/1062-6050-52.6.04
32. Chamorro C, Armijo-Olivo S, De la Fuente C, Fuentes J, Javier Chirisa L. Absolute reliability and concurrent validity of hand held dynamometry and isokinetic dynamometry in the hip, knee and ankle joint: Systematic review and meta-analysis. *Open Med*. 2017;12:359-375. doi:10.1515/med-2017-0052
33. Mentiplay BF, Perraton LG, Bower KJ, et al. Assessment of lower limb muscle strength and power using hand-held and fixed dynamometry: A reliability and validity Study. *PLoS One*. 2015;10(10):e0140822. doi:10.1371/journal.pone.0140822
34. Arnold CM, Warkentin KD, Chilibeck PD, Magnus CR. The reliability and validity of handheld dynamometry for the measurement of lower-extremity muscle strength in older adults. *J Strength Cond Res*. 2010;24(3):815-824. doi:10.1519/JSC.0b013e3181aa36b8
35. Hirano M, Katoh M, Gomi M, Arai S. Validity and reliability of isometric knee extension muscle strength measurements using a belt-stabilized hand-held dynamometer: A comparison with the measurement using an isokinetic dynamometer in a sitting posture. *J Phys Ther Sci*. 2020;32(2):120-124. doi:10.1589/jpts.32.120

36. Lesnak J, Anderson D, Farmer B, Katsavelis D, Grindstaff TL. Validity of hand-held dynamometry in measuring quadriceps strength and rate of torque development. *Int J Sports Phys Ther*. 2019;14(2):180-187. [doi:10.26603/ijsp.20190180](https://doi.org/10.26603/ijsp.20190180)
37. Beuchat A, Maffiuletti NA. Foot rotation influences the activity of medial and lateral hamstrings during conventional rehabilitation exercises in patients following anterior cruciate ligament reconstruction. *Phys Ther Sport*. 2019;39:69-75. [doi:10.1016/j.ptsp.2019.06.010](https://doi.org/10.1016/j.ptsp.2019.06.010)
38. Hogberg J, Bergentoft E, Piusi R, et al. Persistent knee flexor strength deficits identified through the NordBord eccentric test not seen with “gold standard” isokinetic concentric testing during the first year after anterior cruciate ligament reconstruction with a hamstring tendon autograft. *Phys Ther Sport*. 2022;55:119-124. [doi:10.1016/j.ptsp.2022.03.004](https://doi.org/10.1016/j.ptsp.2022.03.004)